



**Intronics  
Power®**

# Economy, Wideband Log/Antilog Amplifiers

**MODEL 759N 759P**

## FEATURES

- Low-Cost, Complete Log/Antilog Amplifier
- External Components Not Required;
- Internal Reference; Temperature Compensated
- Small Size: 1.1" x 1.1" x 0.4"
- Fast Response: 200kHz Bandwidth ( $I_{SIG} = 1\mu A$ )
- 6 Decades Current Operation – 1nA to 1mA
  - 1% max Error – 20nA to 200 $\mu A$
  - 2% max Error – 10nA to 1mA
- 4 Decades Voltages Operation – 1mV to 10V
  - 1% max Error – 1mV to 2V
  - 2% max Error – 1mV to 10V

## APPLICATIONS

- Log Current or Voltage
- Antilog Voltage
- Data Compression or Expansion



## GENERAL DESCRIPTION

Models 759N and 759P are low cost, fast response, dc logarithmic amplifiers offering 1% conformance to ideal log operation over four decades of current operation – 20nA to 200 $\mu A$ , as well as 2% conformance over four decades of voltage operation – 1mV to 10V. Featuring 200kHz bandwidth at  $I_{SIG} = 1\mu A$ , these new economy designs are the industry's fastest log/antilog amplifiers and offer an attractive alternative to in-house designs.

Designed for ease of use, models 759N/P are complete, temperature compensated, log or antilog amplifiers packaged in a small 1.1" x 1.1" x 0.4" epoxy encapsulated module. External components are not required for logging currents over the complete six decade operating range from 1nA to 1mA. Both the scale factor ( $K = 2, 1, 2/3$  volt/decade) and log/antilog operation can be selected by simple pin interconnection. In addition both the internal 10 $\mu A$  reference current as well as the offset voltage may be externally adjusted to improve overall accuracy performance.

## MODEL SELECTION

Model 759N computes the log of positive input signals (voltage or current), while model 759P computes the log of negative input signals (voltage or current). In the antilog mode of operation, both models accept bipolar voltage input signals ( $-2V \leq E_{SIG} / K \leq 2V$ ), with model 759N producing a positive output signal and model 759P producing a negative output signal.

## APPLICATIONS

Model 759N and 759P can operate with either current or voltage inputs when connected as shown in Figure 1. To illustrate the logarithmic transfer characteristics, a plot of input current versus output voltage is also presented. Model 759 is ideally

suitable for log applications whenever low cost implementation of logarithmic natural relationships is advantageous. Examples are absorbance measurements, data compression and expansion, chemical analysis of liquids, computing powers, roots and ratios and conversion of exponential quantities to linear form.





Figure 1. Functional Block Diagram and Transfer Function

# SPECIFICATIONS

(typical @ +25°C and  $V_S = \pm 15V$  dc unless otherwise noted)

MODEL	759N/P		
<b>TRANSFER FUNCTIONS</b>			
Current Mode	$e_O = -K \log_{10} \frac{I_{SIG}}{I_{REF}}$		
Voltage Mode	$e_O = -K \log_{10} \frac{E_{SIG}}{E_{REF}}$		
Antilog Mode	$e_O = E_{REF} 10^{\left(\frac{E_{SIG}}{K}\right)}$		
<b>TRANSFER FUNCTION PARAMETERS</b>			
Scale Factor (K) Selections <sup>1, 2</sup>	2, 1, 2/3 Volt/Decade		
Error @ +25°C	±1% max		
vs. Temperature (0 to +70°C)	±0.04%/°C max		
Reference Voltage ( $E_{REF}$ ) <sup>2</sup>	0.1V		
Error @ +25°C	±4% max		
vs. Temperature (0 to +70°C)	±0.05%/°C		
Reference Current ( $I_{REF}$ ) <sup>2</sup>	10μA		
Error @ +25°C	±3% max		
vs. Temperature (0 to +70°C)	±0.05%/°C		
<b>LOG CONFORMITY ERROR</b>			
$I_{SIG}$ Range	$E_{SIG}$ Range	R.T.I.	R.T.O. (K = 1)
20nA to 200μA	1mV to 2V	±1% max	±4.3mV max
1nA to 1mA	1mV to 10V	±2% max	±8.64mV max
1nA to 10nA		±5%	±21mV
<b>INPUT SPECIFICATIONS</b>			
Current Signal Range			
Model 759N	+1nA to +1mA min		
Model 759P	-1nA to -1mA min		
Max Safe Input Current	±10mA max		
Bias Current @ +25°C	(0, +) 200pA max		
vs. Temperature (0 to +70°C)	x2/+10°C		
Voltage Signal Range			
Model 759N	+1mV to +10V min		
Model 759P	-1mV to -10V min		
Offset Voltage @ +25°C (Adjustable to 0)	±2mV max		
vs. Temperature (0 to +70°C)	±10μV/°C		
vs. Supply Voltage	±15μV/%		
<b>FREQUENCY RESPONSE, Sinewave</b>			
Small Signal Bandwidth, -3dB			
$I_{SIG} = 1nA$	250Hz		
$I_{SIG} = 10nA$	1.8kHz		
$I_{SIG} = 100nA$	25kHz		
$I_{SIG} = 1μA$	200kHz		
$I_{SIG} = 10μA$	300kHz		
$I_{SIG} = 100μA$	300kHz		
$I_{SIG} = 1mA$	300kHz		
<b>RISE TIME</b>			
Increasing Input Current			
10nA to 100nA	20μs		
100nA to 1μA	3μs		
1μA to 100μA	2.5μs		
100μA to 1mA	2.5μs		
Decreasing Input Current			
1mA to 100μA	3μs		
100μA to 1μA	3μs		
1μA to 100nA	10μs		
100nA to 10nA	80μs		
<b>INPUT NOISE</b>			
Voltage, 10Hz to 10kHz	10μV rms		
Current, 10Hz to 10kHz	10pA rms		
<b>OUTPUT SPECIFICATIONS<sup>3</sup></b>			
Rated Output			
Voltage	±10V min		
Current			
Log Mode	±5mA		
Antilog Mode	±4mA		
Resistance	0.5Ω		
<b>POWER SUPPLY<sup>4</sup></b>			
Rated Performance	±15V dc		
Operating	±(12 to 18)V dc		
Current, Quiescent	±4mA		
<b>TEMPERATURE RANGE</b>			
Rated Performance	0 to +70°C		
Operating	-25°C to +85°C		
Storage	-55°C to +125°C		
CASE SIZE	1.125" x 1.125" x 0.4"		

<sup>1</sup> Use terminal 1 for K = 1V/decade, terminal 2 for K = 2V/decade, terminals 1 and 2 (shorted together) for K = 2/3V/decade.

<sup>2</sup> Specification is + for model 759N, - for model 759P.


<sup>3</sup> No damage due to any pin being shorted to ground.

<sup>4</sup> Recommended power supply, model 904, ±15V @ ±50mA output.

Specifications subject to change without notice.

## OUTLINE DIMENSIONS

Dimensions shown in inches and (mm).




WEIGHT: 10 GRAMS


MATING SOCKET AC1016 \$5.00 (1-9)

<sup>1</sup> Optional 100kΩ external trim pot. Input offset voltage may be adjusted to zero with trim pot connected as shown. With trim terminal 9 left open, input offset voltage will be ±2mV maximum.

## TRANSFER CURVES



Plot of Output Voltage vs Input Current for Model 759 Connected in the Log Mode



Plot of Output Voltage vs Input Voltage for Model 759 Connected in the Log Mode

# Understanding the Log Amplifier Performance

## PRINCIPLE OF OPERATION

Log operation is obtained by placing the antilog element in the feedback loop of the op amp as shown in Figure 1. At the summing junction, terminal 5, the input signal current to be processed is summed with the output current of the antilog element. To attain a balance of these two currents, the op amp provides the required output voltage to the antilog feedback element. Under these conditions the ideal transfer equation ( $K = 1$ ) is:

$$e_{OUT} = 1V \log_{10} I_{SIG}/I_{REF}$$

The log is a mathematical operator which is defined only for numbers, which are dimensionless quantities. Since an input current would have the dimensions of amperes it must be referenced to another current,  $I_{REF}$ , the ratio being dimensionless. For this purpose a temperature compensated reference of  $10\mu A$  is generated internally.

The scale factor,  $K$ , is a multiplying constant. For a change in input current of one decade (decade = ratio of 10:1), the output changes by  $K$  volts.  $K$  may be selected as 1V or 2V by connecting the output to pin 1 or 2, respectively. If the output is connected to both pins 1 and 2,  $K$  will be 2/3V.

## REFERRING ERRORS TO INPUT

A unique property of log amplifiers is that a dc error of any given amount at the output corresponds to a constant percent of the input, regardless of input level. To illustrate this, consider the output effects due to changing the input by 1%.

The output would be:

$$e_{OUT} = 1V \log_{10} (I_{SIG}/I_{REF})(1.01) \text{ which is equivalent to:}$$

$$e_{OUT} = \underbrace{1V \log_{10} (I_{SIG}/I_{REF})}_{\text{Initial Value}} \underbrace{\pm 1V \log_{10} (1.01)}_{\text{Change}}$$

The change in output, due to a 1% input change is a constant value of  $\pm 4.3mV$ . Conversely, a dc error at the output of  $\pm 4.3mV$  is equivalent to a change at the input of 1%. An abbreviated table is presented below for converting between errors referred to output (R.T.O.), and errors referred to input (R.T.I.).

ERROR R.T.I.	ERROR R.T.O.		
	K = 1	K = 2	K = 2/3
0.1%	0.43mV	0.86mV	0.28mV
0.5	2.17	4.34	1.45
1.0	4.32	8.64	2.88
3.0	12.84	25.68	8.56
4.0	17.03	34.06	11.35
5.0	21.19	42.38	14.13
10.0	41.39	82.78	27.59

Table 1. Converting Output Error in mV to Input Error in %

## SOURCES OF ERROR

**Log Conformity Error** – Log conformity in logarithmic devices is a specification similar to linearity in linear devices. Log conformity error is the difference between the value of the transfer equation and the actual value which occurs at the output of the log module, after scale factor, reference and offset errors are eliminated or taken into account. Figure 2 below illustrates the log conformity performance of model 759 over a 6 decade input range. The best linearity performance is obtained in the 5 decades from 10nA to 1mA. To obtain optimum performance, the input data should be scaled to this range.




Figure 2. Log Conformity Error for Models 759N and 759P

**Offset Voltage** – The offset voltage,  $E_{OS}$ , of model 759 is the offset voltage of the internal FET amplifier. This voltage appears as a small dc offset voltage in series with the input terminals. For current logging applications, its error contribution is negligible. However, for log voltage applications, best performance is obtained by an offset trim adjustment.

**Bias Current** – The bias current of model 759 is the bias current of the internal FET amplifier. This parameter can be a significant source of error when processing signals in the nanoamp region. For this reason, the bias current for model 759 is 200pA, maximum.

**Reference Current** –  $I_{REF}$  is the internally generated current source to which all input currents are compared.  $I_{REF}$  tolerance errors appear as a dc offset at the output. The specified value of  $I_{REF}$  is  $\pm 3\%$  referred to the input, and, from Table 1, corresponds to a dc offset of  $\pm 12.84mV$  for  $K = 1$ . This offset is independent of input signal and may be removed by injecting a current into terminal 1 or 2.

**Reference Voltage** –  $E_{REF}$  is the effective internally generated voltage to which all input voltages are compared. It is related to  $I_{REF}$  through the equation:  $E_{REF} = I_{REF} \times R_{IN}$ , where  $R_{IN}$  is an internal  $10k\Omega$ , precision resistor. Virtually all tolerance in  $E_{REF}$  is due to  $I_{REF}$ . Consequently, variations in  $I_{REF}$  cause a shift in  $E_{REF}$ .

**Scale Factor** – Scale factor is the voltage change at the output for a decade (i.e., 10:1) change at the input, when connected in the log mode. Error in scale factor is equivalent to a change in gain, or slope, and is specified in per cent of the nominal value. An external adjustment may be performed if fine trimming is desired for improved accuracy.

## OPTIONAL EXTERNAL ADJUSTMENTS FOR LOG OPERATION

**Trimming  $E_{OS}$**  – The amplifier's offset voltage,  $E_{OS}$ , may be trimmed for improved accuracy with the model 759 connected in its log circuit. To accomplish this, a 100k $\Omega$ , 10 turn pot is connected as shown in Figure 3. The input terminal, Pin 4, is connected to ground. Under these conditions the output voltage is:

$$e_{OUT} = -K \log_{10} E_{OS}/E_{REF}$$

To obtain an offset voltage of 100 $\mu$ V or less, for  $K = 1$ , the trim pot should be adjusted until the output voltage is between +3 and +4 volts for model 759N, and -3 to -4V for model 759P.

For other values of  $K$ , the trim pot should be adjusted for an output of  $e_{OUT} = 3 \times K$  to  $4 \times K$  where  $K$  is the scale factor.




Figure 3. Trimming  $E_{OS}$  in Log Mode

**Reference Current or Reference Voltage** – The reference current or voltage of model 759 may be shifted by injecting a constant current into the unused scale factor terminal (Pin 1 or Pin 2). Each 330 $\mu$ A of current injected will shift the reference one decade, in accordance with the expression:  $I_1 = 330\mu A \log 10\mu A/I_{REF}$ , where  $I_1$  = current to be injected and  $I_{REF}$  = the desired reference current.

By changing  $I_{REF}$ , there is a corresponding change in  $E_{REF}$  since,  $E_{REF} = I_{REF} \times R_{IN}$ . An alternate method for rescaling  $E_{REF}$  is to connect an external  $R_{IN}$ , at the  $I_{IN}$  terminal (Pin 5) to supplant the 10k $\Omega$  supplied internally (leaving it unconnected). The expression for  $E_{REF}$  is then,  $E_{REF} = R_{IN} I_{REF}$ . Care must be taken to choose  $R_{IN}$  such that  $(e_{SIG} \text{ max})/R_{IN} \leq 1\text{mA}$ .

**Scale Factor (K) Adjustment** – Scale factor may be increased from its nominal value by inserting a series resistor between the output terminal, Pin 3, and either terminal 1 or 2. The table below should be consulted when making these scale factor changes.

RANGE OF K	CONNECT SERIES R TO PIN	VALUE OF R	NOTE
2/3V to 1.01V	1	3k $\Omega$ x (K - 2/3)	use pins 1, 2
1.01V to 2.02V	1	3k $\Omega$ x (K - 1)	use pin 1
>2.02V	2	3k $\Omega$ x (K - 2)	use pin 2

Table 2. Resistor Selection Chart for Shifting Scale Factor

### ANTILOG OPERATION

The model 759 may be used to develop the antilog of the input voltage when connected as shown in Figure 4. The antilog transfer function (an exponential), is:

$$e_{OUT} = E_{REF} 10^{-e_{IN}/K} \quad [-2 \leq e_{IN}/K \leq 2]$$




Figure 4. Functional Block Diagram

**Principle of Operation** – The antilog element converts the voltage input, appearing at terminal 1 or 2, to a current which is proportional to the antilog of the applied voltage. The current-to-voltage conversion is then completed by the feedback resistor in a closed-loop op amp circuit.

A more complete expression for the antilog function is:

$$e_{OUT} = E_{REF} 10^{-e_{IN}/K} + E_{OS}$$

The terms  $K$ ,  $E_{OS}$ , and  $E_{REF}$  are those described previously in the LOG section.

**Offset Voltage ( $E_{OS}$ ) Adjustment** – Although offset voltage of the antilog circuit may be balanced by connecting it in the log mode, and using the technique described previously, it may be more advantageous to use the circuit of Figure 5. In this configuration, offset voltage is equal to  $e_{OUT}/100$ . Adjust for the desired null, using the 100k trim pot. After adjusting, turn power off, remove the external 100 $\Omega$  resistor, and the jumper from Pin 1 to +15V. For 759P use the same procedure but connect Pin 1 to -15V.




Figure 5. Trimming  $E_{OS}$  in Antilog Mode

**Reference Voltage ( $E_{REF}$ ) Adjustment** – In antilog operation, the voltage reference appears as a multiplying constant.  $E_{REF}$  adjustment may be accomplished by connecting a resistor,  $R$ , from Pin 5 to Pin 3, in place of the internal 10k $\Omega$ . The value of  $R$  is determined by:

$$R = E_{REF} \text{ desired}/10^{-5} \text{ A}$$

**Scale Factor (K) Adjustment** – The scale factor may be adjusted for all values of  $K$  greater than 2/3V by the techniques described in the log section. If a value of  $K$ , less than 2/3V is desired for a given application, an external op amp would be required as shown in Figure 6. The ratio of the two resistors is approximately:

$$R_1/R_G = (1/K - 1) \text{ where } K = \text{desired scale factor}$$




Figure 6. Method for Adjusting  $K < 2/3V$